



HSG Catheter Set available in 5(fr) and 7(fr) sizes

HSG catheter sets can be used for sonohysterography and hysterosalpingography procedures that greatly enhance the analysis of uterine pathology. Using HSG balloon catheters, the procedures are simple and non-traumatic for your patients. A placement sheath eliminates the need for a tenaculum. HSG catheters are designed for single patient use only.

Product No.: MM1154F (5 French Catheter)FIRM
MM1155F (7 French Catheter)FIRM
MM1154 (5 French Catheter)
MM1155 (7 French Catheter)

CAUTION:

Federal law requires that these devices be ordered by a physician.

INDICATIONS:

HSG Catheter Sets are indicated to evaluate the causes of abnormal uterine bleeding, menstrual disorders, recurring pregnancy loss, or unexplained infertility. They can also be used to assess uterine pathology and patients on tamoxifen therapy. HSG type catheter sets are used to infuse a fluid (either a contrast media or a sterile saline) into the uterine cavity while blocking the external cervical OS to retain the fluid in the uterus during the procedure.

CONTRAINDICATIONS:

HSG procedures are contraindicated in pregnancies, suspected pregnancies, active pelvic infections, or recent pelvic infections, severe uterine bleeding, gynecologic malignancies, or if the patient is allergic to the contrast media. This device should not be used for intrafallopian procedures.

PRECAUTIONS:

Do not exceed the preset recommended balloon inflation volume (2 cc for the 5 French catheter and 3.5 cc for the 7 French catheter) or the balloon may burst. The use of OIL-BASED contrast media such as ethyl esters may interact with the balloon of the catheter,

Distributed by:
Monarch Medical Products
monarchmedicalproducts.com
Toll Free Phone: 866-241-1625

NOT MADE WITH NATURAL RUBBER LATEX

causing possible balloon rupture. The use of aqueous contrast media is recommended.

PROCEDURES:

Remove the HSG set from the sterile pouch and remove the balloon protective sleeve from the distal end of the catheter. Using the attached syringe, inflate the silicone balloon by opening the stopcock attached to the syringe luer. Then allow the silicone balloon to completely deflate. This procedure ensures the integrity of the HSG balloon. Attach the media source to the open luer and allow the contrast media to flow through the catheter removing air and clearing the media channel.

DIRECTIONS FOR USE:

With the guiding sheath, position the HSG catheter so that the tip of the catheter is at the entrance of the cervix external OS. Gently push the HSG catheter, with balloon, through the cervical canal and into the uterine cavity. Inflate the balloon within the uterine cavity and close the balloon stopcock to prevent the balloon from deflating. Carefully withdraw the catheter until the balloon is resting against the internal OS. Now inject the contrast media into the uterus and proceed with the examination.

In cases where the access to the uterus is difficult, the balloon may be inflated within the cervical canal.



DO NOT REUSE



CAUTION



USE BY



BATCH CODE



STERILE EO